

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051224.D
 Acq On : 15 Sep 2018 15:55
 Operator : MD\SY
 Sample : PB112879ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#05

Quant Time: Sep 17 03:29:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	428015	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	682468	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	539581	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	161083	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	232936	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	7.59	113	250082	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.09	98	851546	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
62) 4-Bromofluorobenzene	12.40	95	218447	33.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.88%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	8258	1.20	ug/l	93
18) Methyl Acetate	4.34	43	11432	3.00	ug/l #	90
20) Methylene Chloride	4.55	84	5946	1.13	ug/l	97
43) Isopropyl Acetate	8.17	43	234023	30.56	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051224.D
Acq On : 15 Sep 2018 15:55
Operator : MD\SY
Sample : PB112879ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#05

Quant Time: Sep 17 03:29:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

