

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051232.D
 Acq On : 15 Sep 2018 19:11
 Operator : MD\SY
 Sample : PB112879ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#13

Quant Time: Sep 17 04:06:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	417473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	663128	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	552707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	164910	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	232920	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	7.59	113	248434	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	10.09	98	849313	44.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.84%	
62) 4-Bromofluorobenzene	12.40	95	224449	35.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.78%	
Target Compounds						
16) Acetone	3.82	43	8783	1.85	ug/l	93
18) Methyl Acetate	4.33	43	12908	3.47	ug/l	99
20) Methylene Chloride	4.56	84	6055	1.18	ug/l	88
43) Isopropyl Acetate	8.17	43	259822	34.92	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051232.D
Acq On : 15 Sep 2018 19:11
Operator : MD\SY
Sample : PB112879ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#13

Quant Time: Sep 17 04:06:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

