

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051233.D
 Acq On : 15 Sep 2018 19:36
 Operator : MD\SY
 Sample : PB112879ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#14

Quant Time: Sep 17 04:08:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	416229	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	665247	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	535167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	154913	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	229849	49.62	ug/l	0.00
Spiked Amount			Recovery	=	99.24%	
35) Dibromofluoromethane	7.59	113	247572	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
50) Toluene-d8	10.09	98	844793	44.54	ug/l	0.00
Spiked Amount			Recovery	=	89.08%	
62) 4-Bromofluorobenzene	12.40	95	213095	33.96	ug/l	0.00
Spiked Amount			Recovery	=	67.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	9416	2.44	ug/l	93
18) Methyl Acetate	4.34	43	13577	3.66	ug/l	96
20) Methylene Chloride	4.55	84	7791	1.52	ug/l	91
43) Isopropyl Acetate	8.17	43	246289	33.00	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051233.D
Acq On : 15 Sep 2018 19:36
Operator : MD\SY
Sample : PB112879ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#14

Quant Time: Sep 17 04:08:09 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

