

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051235.D
 Acq On : 15 Sep 2018 20:25
 Operator : MD\SY
 Sample : PB112879ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#16

Quant Time: Sep 17 04:14:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	402860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	653827	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	522022	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	152129	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	226348	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
35) Dibromofluoromethane	7.59	113	239239	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.09	98	816694	43.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.62%	
62) 4-Bromofluorobenzene	12.40	95	210409	34.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.24%	
Target Compounds						
16) Acetone	3.82	43	8893	2.24	ug/l	95
18) Methyl Acetate	4.33	43	15716	4.38	ug/l	96
20) Methylene Chloride	4.55	84	8209	1.66	ug/l	94
43) Isopropyl Acetate	8.17	43	299216	40.79	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051235.D
Acq On : 15 Sep 2018 20:25
Operator : MD\SY
Sample : PB112879ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#16

Quant Time: Sep 17 04:14:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

