

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051236.D
 Acq On : 15 Sep 2018 20:50
 Operator : MD\SY
 Sample : PB112879ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#17

Quant Time: Sep 17 04:17:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	420047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	670975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	537317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	162756	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	230833	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.59	113	246857	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	10.09	98	843599	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
62) 4-Bromofluorobenzene	12.40	95	217889	34.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	8650	1.69	ug/l	92
18) Methyl Acetate	4.33	43	12094	3.23	ug/l	98
20) Methylene Chloride	4.55	84	12438	2.41	ug/l	94
43) Isopropyl Acetate	8.17	43	245239	32.58	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051236.D
Acq On : 15 Sep 2018 20:50
Operator : MD\SY
Sample : PB112879ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#17

Quant Time: Sep 17 04:17:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

