

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051246.D
 Acq On : 16 Sep 2018 1:44
 Operator : MD\SY
 Sample : J4773-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-091418-B

Quant Time: Sep 17 05:26:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	423717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	669656	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	538770	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	168038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	236337	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
35) Dibromofluoromethane	7.59	113	250502	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
50) Toluene-d8	10.09	98	845866	44.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.60%	
62) 4-Bromofluorobenzene	12.40	95	221281	35.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.06%	
Target Compounds						
16) Acetone	3.82	43	17145	9.08	ug/l	96
18) Methyl Acetate	4.33	43	9380	2.48	ug/l #	88
20) Methylene Chloride	4.55	84	362337	69.56	ug/l	99
43) Isopropyl Acetate	8.17	43	182411	24.28	ug/l #	84
95) Naphthalene	15.13	128	22437	8.51	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051246.D
Acq On : 16 Sep 2018 1:44
Operator : MD\SY
Sample : J4773-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-091418-B

Quant Time: Sep 17 05:26:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

