

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051251.D
 Acq On : 16 Sep 2018 3:46
 Operator : MD\SY
 Sample : J4929-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-091418-B

Quant Time: Sep 17 05:36:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	410825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	651266	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	516064	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	152877	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	227768	49.82	ug/l	0.00
Spiked Amount			Recovery	=	99.64%	
35) Dibromofluoromethane	7.59	113	246506	50.37	ug/l	0.00
Spiked Amount			Recovery	=	100.74%	
50) Toluene-d8	10.09	98	829844	44.69	ug/l	0.00
Spiked Amount			Recovery	=	89.38%	
62) 4-Bromofluorobenzene	12.40	95	210339	34.24	ug/l	0.00
Spiked Amount			Recovery	=	68.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	11417	4.36	ug/l	99
18) Methyl Acetate	4.34	43	12236	3.34	ug/l	93
20) Methylene Chloride	4.56	84	266802	52.83	ug/l	99
30) Chloroform	7.37	83	17279	2.07	ug/l	96
43) Isopropyl Acetate	8.17	43	209570	28.68	ug/l #	80
84) 1,2,4-Trimethylbenzene	13.04	105	14210	1.39	ug/l	99
95) Naphthalene	15.13	128	28589	9.78	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051251.D
Acq On : 16 Sep 2018 3:46
Operator : MD\SY
Sample : J4929-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-091418-B

Quant Time: Sep 17 05:36:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

