

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051252.D
 Acq On : 16 Sep 2018 4:11
 Operator : MD\SY
 Sample : J4775-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-091418-A

Quant Time: Sep 17 05:38:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	422788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	665053	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	536053	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	168051	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	232527	49.42	ug/l	0.00
Spiked Amount						
			Recovery	=	98.84%	
35) Dibromofluoromethane	7.59	113	249836	49.99	ug/l	0.00
Spiked Amount						
			Recovery	=	99.98%	
50) Toluene-d8	10.09	98	848661	44.75	ug/l	0.00
Spiked Amount						
			Recovery	=	89.50%	
62) 4-Bromofluorobenzene	12.40	95	218575	34.84	ug/l	0.00
Spiked Amount						
			Recovery	=	69.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	19798	11.45	ug/l	97
18) Methyl Acetate	4.33	43	12757	3.39	ug/l	95
20) Methylene Chloride	4.56	84	45516	8.76	ug/l	97
30) Chloroform	7.38	83	13184	1.53	ug/l	97
43) Isopropyl Acetate	8.17	43	212613	28.49	ug/l #	80
84) 1,2,4-Trimethylbenzene	13.04	105	13102	1.17	ug/l	98
95) Naphthalene	15.13	128	13947	7.31	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051252.D
Acq On : 16 Sep 2018 4:11
Operator : MD\SY
Sample : J4775-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-091418-A

Quant Time: Sep 17 05:38:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

