

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051253.D
 Acq On : 16 Sep 2018 4:36
 Operator : MD\SY
 Sample : J4774-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-091418

Manual Integrations
 APPROVED

MMDadoda
 9/18/2018 4:32:12 PM

Quant Time: Sep 17 05:40:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	430545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	669393	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	543239	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	162061	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	237094	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	7.59	113	255414	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
50) Toluene-d8	10.09	98	854749	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
62) 4-Bromofluorobenzene	12.40	95	222330	35.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.42%	
Target Compounds						
16) Acetone	3.83	43	11432	3.90	ug/l	96
18) Methyl Acetate	4.34	43	11483	2.99	ug/l #	77
20) Methylene Chloride	4.55	84	13261m	2.51	ug/l	
43) Isopropyl Acetate	8.17	43	188812	25.14	ug/l #	80
95) Naphthalene	15.13	128	12079	7.11	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051253.D
Acq On : 16 Sep 2018 4:36
Operator : MD\SY
Sample : J4774-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-01-091418

Manual Integrations
APPROVED

MMDadoda
9/18/2018 4:32:12 PM

Quant Time: Sep 17 05:40:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

