

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051254.D
 Acq On : 16 Sep 2018 5:00
 Operator : MD\SY
 Sample : J4935-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PHASE-1

Quant Time: Sep 17 05:43:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	436060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	676998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	546716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	174786	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	240084	49.47	ug/l	0.00
Spiked Amount						
			Recovery	=	98.94%	
35) Dibromofluoromethane	7.59	113	255239	50.17	ug/l	0.00
Spiked Amount						
			Recovery	=	100.34%	
50) Toluene-d8	10.09	98	863794	44.75	ug/l	0.00
Spiked Amount						
			Recovery	=	89.50%	
62) 4-Bromofluorobenzene	12.40	95	226583	35.48	ug/l	0.00
Spiked Amount						
			Recovery	=	70.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	10353	2.86	ug/l	99
18) Methyl Acetate	4.34	43	11695	3.01	ug/l	99
20) Methylene Chloride	4.55	84	6155638	1148.36	ug/l	100
30) Chloroform	7.37	83	16346	1.84	ug/l	100
43) Isopropyl Acetate	8.17	43	200436	26.39	ug/l #	82
95) Naphthalene	15.13	128	19304	7.96	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051254.D
Acq On : 16 Sep 2018 5:00
Operator : MD\SY
Sample : J4935-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PHASE-1

Quant Time: Sep 17 05:43:20 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

