

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051256.D
 Acq On : 16 Sep 2018 5:49
 Operator : MD\SY
 Sample : J4773-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-091418-D

Quant Time: Sep 17 05:46:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	413117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	645046	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	516191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	152730	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	224381	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
35) Dibromofluoromethane	7.59	113	241409	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
50) Toluene-d8	10.09	98	823550	44.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.56%	
62) 4-Bromofluorobenzene	12.40	95	207028	34.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	68.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	21138	13.06	ug/l	89
18) Methyl Acetate	4.34	43	11524	3.13	ug/l	96
20) Methylene Chloride	4.55	84	21196	4.17	ug/l	93
30) Chloroform	7.38	83	11474	1.36	ug/l	88
43) Isopropyl Acetate	8.17	43	185638	25.65	ug/l #	80
95) Naphthalene	15.13	128	16497	7.91	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051256.D
Acq On : 16 Sep 2018 5:49
Operator : MD\SY
Sample : J4773-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-091418-D

Quant Time: Sep 17 05:46:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

