

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051236.D
 Acq On : 15 Sep 2018 20:50
 Operator : MD\SY
 Sample : PB112879ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#17

Quant Time: Sep 17 04:17:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	420047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	670975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	537317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	162756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	230833	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.59	113	246857	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	10.09	98	843599	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
62) 4-Bromofluorobenzene	12.40	95	217889	34.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.86%	
Target Compounds						
16) Acetone	3.82	43	8650	1.69	ug/l	92
18) Methyl Acetate	4.33	43	12094	3.23	ug/l	98
20) Methylene Chloride	4.55	84	12438	2.41	ug/l	94
43) Isopropyl Acetate	8.17	43	245239	32.58	ug/l #	79


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051236.D
Acq On : 15 Sep 2018 20:50
Operator : MD\SY
Sample : PB112879ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#17

Quant Time: Sep 17 04:17:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN051236.D

Acq: 15 Sep 2018 20:50

Instrument :

MSVOA_N

ClientSampleId :


PB112879ZHE#17

Tot Ion:168 Resp: 420047

Ion Ratio Lower Upper

168 100

99 47.6 38.3 57.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

200000



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 1.69 ug/l

RT: 3.82 min Scan# 638

Delta R.T. 0.01 min

Lab File: VN051236.D


Acq: 15 Sep 2018 20:50

Tgt Ion: 43 Resp: 8650

Ion Ratio Lower Upper

43 100

58 30.1 27.8 41.6



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.82


3000


2000

1000

0

Time-->









#18
Methyl Acetate
Concen: 3.23 ug/l
RT: 4.33 min Scan# 796
Delta R.T. 0.01 min
Lab File: VN051236.D
Acq: 15 Sep 2018 20:50




Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#17


Tot Ion: 43 Resp: 12094
Ion Ratio Lower Upper
43 100
74 25.1 20.7 31.1



#20
Methylene Chloride
Concen: 2.41 ug/l
RT: 4.55 min Scan# 864
Delta R.T. -0.00 min
Lab File: VN051236.D
Acq: 15 Sep 2018 20:50

Tgt Ion: 84 Resp: 12438
Ion Ratio Lower Upper
84 100
49 107.2 90.2 135.2
51 32.1 28.4 42.6
86 58.2 51.8 77.8









#33
 1,2-Dichloroethane-d4
 Concen: 49.38 ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN051236.D
 Acq: 15 Sep 2018 20:50




Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#17

Tot Ion: 65 Resp: 230833
 Ion Ratio Lower Upper
 65 100
 67 54.8 0.0 112.0



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: VN051236.D
 Acq: 15 Sep 2018 20:50

Tgt Ion: 114 Resp: 670975
 Ion Ratio Lower Upper
 114 100
 63 19.1 0.0 37.8
 88 15.7 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 48.96 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN051236.D
 Acq: 15 Sep 2018 20:50




Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#17


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.2	83.0	124.6
192	23.7	18.7	28.1



#43
 Isopropyl Acetate
 Concen: 32.58 ug/l
 RT: 8.17 min Scan# 1989
 Delta R.T. 0.01 min
 Lab File: VN051236.D
 Acq: 15 Sep 2018 20:50

Tgt Ion	Ratio	Lower	Upper
43	100		
61	16.4	25.0	37.6#
87	10.7	11.8	17.8#





#50

Toluene-d8

Concen: 44.09 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN051236.D

Acq: 15 Sep 2018 20:50

Instrument :

MSVOA_N

ClientSampleId :





PB112879ZHE#17

Tot Ion: 98 Resp: 843599

Ion Ratio Lower Upper

98 100

100 62.9 52.0 78.0



#62

4-Bromofluorobenzene

Concen: 34.43 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN051236.D

Acq: 15 Sep 2018 20:50




Tgt Ion: 95 Resp: 217889


Ion Ratio Lower Upper

95 100

174 95.6 0.0 184.8

176 91.7 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN051236.D
Acq: 15 Sep 2018 20:50

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#17

Tot Ion:117 Resp: 537317
Ion Ratio Lower Upper
117 100
82 51.7 41.8 62.6
119 31.8 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN051236.D
Acq: 15 Sep 2018 20:50

Tgt Ion:152 Resp: 162756
Ion Ratio Lower Upper
152 100
115 54.4 28.2 84.7
150 158.5 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

