

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091519\
 Data File : VN058097.D
 Acq On : 15 Sep 2019 17:12
 Operator : JC/SP
 Sample : K4719-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :


Quant Time: Sep 16 03:42:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	197013	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	327124	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	306138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	160794	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	356167	89.26	ug/l	0.00
Spiked Amount			Recovery	=	178.52%	
35) Dibromofluoromethane	7.58	113	243527	80.66	ug/l	0.00
Spiked Amount			Recovery	=	161.32%	
50) Toluene-d8	10.09	98	967220	82.51	ug/l	0.00
Spiked Amount			Recovery	=	165.02%	
62) 4-Bromofluorobenzene	12.40	95	324035	72.54	ug/l	0.00
Spiked Amount			Recovery	=	145.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.04	50	675	0.243	ug/l	98
5) Bromomethane	2.46	94	32	Below Cal	#	2
16) Acetone	3.80	43	17564	13.427	ug/l #	84
18) Methyl Acetate	4.30	43	7479	2.144	ug/l	98
20) Methylene Chloride	4.53	84	12069	3.707	ug/l	95
25) 2-Butanone	6.83	43	2471	1.305	ug/l	97
29) Tetrahydrofuran	7.17	42	249	0.213	ug/l #	34
31) Cyclohexane	7.65	56	5255	Below Cal	#	25
36) 1,1-Dichloropropene	7.65	75	16719	4.407	ug/l #	50
38) Carbon Tetrachloride	7.65	117	20016	5.779	ug/l #	18
43) Isopropyl Acetate	8.15	43	133584	20.266	ug/l #	83
45) 1,2-Dichloropropane	9.26	63	1769	0.532	ug/l #	44
48) Methyl methacrylate	9.17	41	6697	2.003	ug/l #	13
49) 1,4-Dioxane	9.27	88	71	1.256	ug/l #	1
51) 4-Methyl-2-Pentanone	9.98	43	10120	2.594	ug/l #	86
54) cis-1,3-Dichloropropene	9.90	75	47654	10.220	ug/l #	70
56) Ethyl methacrylate	10.54	69	2400	0.523	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	5964	1.136	ug/l #	43
59) 2-Hexanone	10.76	43	99948	34.185	ug/l #	52
71) Bromoform	12.40	173	380	0.219	ug/l #	1
74) N-amyl acetate	12.04	43	4963	0.914	ug/l #	49
76) 1,2,3-Trichloropropane	12.41	75	175637	48.286	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.41	75	175637	169.919	ug/l #	6
95) Naphthalene	15.13	128	207	Below Cal	#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 197013

Ion Ratio Lower Upper




168 100

99 59.8 48.0 72.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#3

Chloromethane

Concen: 0.243 ug/l

RT: 2.04 min Scan# 83

Delta R.T. 0.01 min

Lab File: VN058097.D


Acq: 15 Sep 2019 17:12

Tgt Ion: 50 Resp: 675

Ion Ratio Lower Upper



50 100


52 32.8 25.5 38.3



Abundance Ion 49.90 (49.60 to 50.60): VN

Ion 51.90 (51.60 to 52.60): VN









#5
Bromomethane
Concen: Below Cal
RT: 2.46 min Scan# 213
Delta R.T. -0.07 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 94 Resp: 32
Ion Ratio Lower Upper
94 100
96 0.0 77.0 115.4#



#16
Acetone
Concen: 13.427 ug/l
RT: 3.80 min Scan# 631
Delta R.T. 0.01 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Tgt Ion: 43 Resp: 17564
Ion Ratio Lower Upper
43 100
58 36.3 22.2 33.2#





#18

Methyl Acetate

Concen: 2.144 ug/l

RT: 4.30 min Scan# 787

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 43 Resp: 7479




Ion Ratio Lower Upper

43 100

74 21.2 17.9 26.9



Abundance Ion 43.00 (42.70 to 43.70): VNC
Ion 74.00 (73.70 to 74.70): VNC



#20

Methylene Chloride

Concen: 3.707 ug/l

RT: 4.53 min Scan# 857

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Tgt Ion: 84 Resp: 12069


Ion Ratio Lower Upper

84 100



49 132.1 104.8 157.2


51 46.3 33.0 49.4

86 71.6 50.5 75.7



Abundance Ion 83.90 (83.60 to 84.60): VNC
Ion 48.90 (48.60 to 49.60): VNC
Ion 50.90 (50.60 to 51.60): VNC
Ion 85.90 (85.60 to 86.60): VNC









#25
2-Butanone
Concen: 1.305 ug/l
RT: 6.83 min Scan# 1573
Delta R.T. 0.01 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 2471
Ion Ratio Lower Upper
43 100
72 27.4 20.6 30.8



#29
Tetrahydrofuran
Concen: 0.213 ug/l
RT: 7.17 min Scan# 1679
Delta R.T. -0.02 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Tgt Ion: 42 Resp: 249
Ion Ratio Lower Upper
42 100
72 0.0 33.7 50.5#
71 0.0 31.9 47.9#









#31
Cyclohexane
Concen: Below Cal
RT: 7.65 min Scan# 1829
Delta R.T. 0.02 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 56 Resp: 5255
Ion Ratio Lower Upper
56 100
69 122.3 25.9 38.9#
84 118.6 65.6 98.4#




Abundance Ion 56.00 (55.70 to 56.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 84.00 (83.70 to 84.70): VNC






#33
1,2-Dichloroethane-d4
Concen: 89.257 ug/l
RT: 8.01 min Scan# 1941
Delta R.T. 0.00 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Tgt Ion: 65 Resp: 356167
Ion Ratio Lower Upper
65 100
67 51.1 0.0 104.0



Abundance Ion 64.90 (64.60 to 65.60): VNC
Ion 66.90 (66.60 to 67.60): VNC









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2115
 Delta R.T. 0.00 min
 Lab File: VN058097.D
 Acq: 15 Sep 2019 17:12

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
114	100		
63	22.4	0.0	44.0
88	15.4	0.0	31.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 80.659 ug/l
 RT: 7.58 min Scan# 1805
 Delta R.T. 0.00 min
 Lab File: VN058097.D
 Acq: 15 Sep 2019 17:12

Tgt Ion	Ratio	Lower	Upper
113	100		
111	104.3	81.3	121.9
192	18.5	14.6	22.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.407 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. -0.13 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 16719

Ion Ratio Lower Upper

75 100

110 8.0 16.3 48.9#




77 0.0 25.0 37.4#



Abundance Ion 74.90 (74.60 to 75.60): VN058097.D (-1774) (-)

Ion 109.90 (109.60 to 110.60): VN058097.D (-1774) (-)

Ion 76.90 (76.60 to 77.60): VN058097.D (-1774) (-)



#38

Carbon Tetrachloride

Concen: 5.779 ug/l

RT: 7.65 min Scan# 1829

Delta R.T. -0.10 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Tgt Ion: 117 Resp: 20016

Ion Ratio Lower Upper

117 100

119 7.4 76.9 115.3#



121 0.0 23.4 35.2#




Abundance Ion 116.80 (116.50 to 117.50): VN058097.D (-1768) (-)

Ion 118.80 (118.50 to 119.50): VN058097.D (-1768) (-)

Ion 120.80 (120.50 to 121.50): VN058097.D (-1768) (-)









#43
Isopropyl Acetate
Concen: 20.266 ug/l
RT: 8.15 min Scan# 1984
Delta R.T. 0.00 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 43 Resp: 133584
Ion Ratio Lower Upper
43 100
61 14.6 20.1 30.1#
87 8.8 9.5 14.3#




Abundance Ion 43.00 (42.70 to 43.70): VNC
Ion 61.00 (60.70 to 61.70): VNC
Ion 87.00 (86.70 to 87.70): VNC






#45
1,2-Dichloropropane
Concen: 0.532 ug/l
RT: 9.26 min Scan# 2329
Delta R.T. 0.15 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Tgt Ion: 63 Resp: 1769
Ion Ratio Lower Upper
63 100
65 0.0 24.8 37.2#



Abundance Ion 62.90 (62.60 to 63.60): VNC
Ion 64.90 (64.60 to 65.60): VNC









#48
Methyl methacrylate
Concen: 2.003 ug/l
RT: 9.17 min Scan# 2301
Delta R.T. -0.02 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 41 Resp: 6697
Ion Ratio Lower Upper
41 100
69 0.0 62.7 94.1#
39 0.0 52.7 79.1#



#49
1,4-Dioxane
Concen: 1.256 ug/l
RT: 9.27 min Scan# 2332
Delta R.T. 0.08 min
Lab File: VN058097.D
Acq: 15 Sep 2019 17:12

Tgt Ion: 88 Resp: 71
Ion Ratio Lower Upper
88 100
43 2038969.0 28.5 42.7#
58 8921.1 55.8 83.6#





#50

Toluene-d8

Concen: 82.511 ug/l

RT: 10.09 min Scan# 2585

Delta R.T. -0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 98 Resp: 967220

Ion Ratio Lower Upper

98 100

100 63.5 50.3 75.5



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

600000

500000

400000

300000

200000

100000

0



10.09

10.00

10.10

10.20

Time-->



#51

4-Methyl-2-Pentanone

Concen: 2.594 ug/l

RT: 9.98 min Scan# 2553

Delta R.T. 0.00 min

Lab File: VN058097.D


Acq: 15 Sep 2019 17:12

Tgt Ion: 43 Resp: 10120

Ion Ratio Lower Upper

43 100

58 28.9 29.8 44.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

30000

25000

20000

15000

10000

5000


0


9.98

9.95

10.00

Time-->






#54




cis-1,3-Dichloropropene
 Concen: 10.220 ug/l
 RT: 9.90 min Scan# 2527
 Delta R.T. 0.06 min
 Lab File: VN058097.D
 Acq: 15 Sep 2019 17:12

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion: 75 Resp: 47654
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.8 37.2#
 39 45.8 46.2 69.4#




Abundance Ion 74.90 (74.60 to 75.60): VNO
 Ion 76.90 (76.60 to 77.60): VNO
 Ion 39.00 (38.70 to 39.70): VNO





#56


Ethyl methacrylate
 Concen: 0.523 ug/l
 RT: 10.54 min Scan# 2725
 Delta R.T. 0.11 min
 Lab File: VN058097.D
 Acq: 15 Sep 2019 17:12

Tgt Ion: 69 Resp: 2400
 Ion Ratio Lower Upper
 69 100
 41 5281.4 56.7 85.1#
 39 2400.5 36.3 54.5#



Abundance Ion 69.00 (68.70 to 69.70): VNO
 Ion 41.00 (40.70 to 41.70): VNO
 Ion 39.00 (38.70 to 39.70): VNO





#57

1,3-Dichloropropane

Concen: 1.136 ug/l

RT: 10.76 min Scan# 2795

Delta R.T. 0.05 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 76 Resp: 5964

Ion Ratio Lower Upper

76 100

78 0.0 25.4 38.2#



#59

2-Hexanone

Concen: 34.185 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.01 min

Lab File: VN058097.D




Acq: 15 Sep 2019 17:12


Tgt Ion: 43 Resp: 99948

Ion Ratio Lower Upper

43 100

58 17.9 25.6 76.8#





#62

4-Bromofluorobenzene

Concen: 72.536 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 95 Resp: 324035

Ion Ratio Lower Upper

95 100

174 74.3 0.0 150.2




176 72.7 0.0 146.0



Abundance Ion 94.90 (94.60 to 95.60): VN058097.D (-2903) (-)

Ion 173.80 (173.50 to 174.50): VN058097.D (-2903) (-)

Ion 175.80 (175.50 to 176.50): VN058097.D (-2903) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Tgt Ion: 117 Resp: 306138

Ion Ratio Lower Upper

117 100

82 58.4 47.2 70.8



119 31.5 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): VN058097.D (-2903) (-)

Ion 82.00 (81.70 to 82.70): VN058097.D (-2903) (-)

Ion 118.90 (118.60 to 119.60): VN058097.D (-2903) (-)





#71

Bromoform

Concen: 0.219 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.27 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:173 Resp: 380

Ion Ratio Lower Upper

173 100

175 4065.8 23.5 70.6#




254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Tgt Ion:152 Resp: 160794

Ion Ratio Lower Upper

152 100

115 61.4 30.3 91.0



150 157.6 0.0 352.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 0.914 ug/l

RT: 12.04 min Scan# 3192

Delta R.T. -0.04 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion: 43 Resp: 4963



Ion Ratio Lower Upper

43 100

70 0.0 30.6 45.8#

55 5.1 18.6 27.8#

61 0.0 17.4 26.2#





Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 48.286 ug/l

RT: 12.41 min Scan# 3307

Delta R.T. -0.15 min

Lab File: VN058097.D



Acq: 15 Sep 2019 17:12

Tgt Ion: 75 Resp: 175637

Ion Ratio Lower Upper


75 100


77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#81

trans-1,4-Dichloro-2-butene

Concen: 169.919 ug/l

RT: 12.41 min Scan# 3307

Delta R.T. 0.10 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 175637

Ion Ratio Lower Upper

75 100

53 0.0 88.4 132.6#




89 0.0 35.3 52.9#



Abundance Ion 74.90 (74.60 to 75.60): VN058097.D (-3307) (-)

Ion 53.00 (52.70 to 53.70): VN058097.D (-3307) (-)

Ion 88.90 (88.60 to 89.60): VN058097.D (-3307) (-)



#95

Naphthalene

Concen: Below Cal

RT: 15.13 min Scan# 4154

Delta R.T. 0.00 min

Lab File: VN058097.D

Acq: 15 Sep 2019 17:12


Tgt Ion: 128 Resp: 207

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#

129 0.0 8.6 12.8#



Abundance Ion 128.00 (127.70 to 128.70): VN058097.D (-4154) (-)

Ion 127.00 (126.70 to 127.70): VN058097.D (-4154) (-)

Ion 129.00 (128.70 to 129.70): VN058097.D (-4154) (-)

