

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091519\
 Data File : VN058091.D
 Acq On : 15 Sep 2019 15:00
 Operator : JC/SP
 Sample : K4695-25
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T5-3-(0)

Quant Time: Sep 16 05:15:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	324588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	546587	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	502436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	268056	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	370455	56.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.70%	
35) Dibromofluoromethane	7.58	113	258261	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
50) Toluene-d8	10.08	98	1001320	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	12.41	95	338398	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
16) Acetone	3.80	43	13719	6.366	ug/l	Qvalue 100
20) Methylene Chloride	4.53	84	10116	1.266	ug/l	94
43) Isopropyl Acetate	8.16	43	168537	15.302	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091519\
Data File : VN058091.D
Acq On : 15 Sep 2019 15:00
Operator : JC/SP
Sample : K4695-25
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T5-3-(0)

Quant Time: Sep 16 05:15:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

