

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091519\
 Data File : VN058097.D
 Acq On : 15 Sep 2019 17:12
 Operator : JC/SP
 Sample : K4719-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-2(0)-BRUCKNER

Quant Time: Sep 16 05:01:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	197013	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	327124	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	306138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	160794	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	356167	89.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	178.52%	
35) Dibromofluoromethane	7.58	113	243527	80.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	161.32%	
50) Toluene-d8	10.09	98	967220	82.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	165.02%	
62) 4-Bromofluorobenzene	12.40	95	324035	72.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	145.08%	
Target Compounds						
16) Acetone	3.80	43	17564	13.427	ug/l #	84
18) Methyl Acetate	4.30	43	7479	2.144	ug/l	98
20) Methylene Chloride	4.53	84	12069	3.707	ug/l	95
43) Isopropyl Acetate	8.15	43	133584	20.266	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091519\
Data File : VN058097.D
Acq On : 15 Sep 2019 17:12
Operator : JC/SP
Sample : K4719-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-2-(0)-BRUCKNER

Quant Time: Sep 16 05:01:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

