

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068480.D
 Acq On : 15 Sep 2021 13:37
 Operator : JC/MD
 Sample : M3719-01
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-091021

Quant Time: Sep 15 16:39:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	521058	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.965	114	750447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	668284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	296734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	247144	51.707	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 103.420%		
35) Dibromofluoromethane	8.019	113	224099	49.512	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 99.020%		
50) Toluene-d8	10.440	98	866796	52.721	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 105.440%		
62) 4-Bromofluorobenzene	12.733	95	276826	52.619	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 105.240%		
Target Compounds						
					Qvalue	
74) N-amyl acetate	12.393	43	6033	1.281	ug/l	95
93) 1,2,4-Trichlorobenzene	15.268	180	8309	1.862	ug/l	94
96) 1,2,3-Trichlorobenzene	15.702	180	9969	2.340	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
Data File : VN068480.D
Acq On : 15 Sep 2021 13:37
Operator : JC/MD
Sample : M3719-01
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-01-091021

Quant Time: Sep 15 16:39:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

