

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068482.D
 Acq On : 15 Sep 2021 14:27
 Operator : JC/MD
 Sample : M3687-21
 Misc : 5.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A005015

Quant Time: Sep 16 03:14:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	528739	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.965	114	766141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	691821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	293062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	259761	53.558	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.120%		
35) Dibromofluoromethane	8.019	113	232107	50.231	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.460%		
50) Toluene-d8	10.443	98	919728	54.795	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 109.580%		
62) 4-Bromofluorobenzene	12.734	95	281052	52.328	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 104.660%		
Target Compounds						
32) 1,1,1-Trichloroethane	8.003	97	9165	1.079	ug/l	Qvalue # 48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
Data File : VN068482.D
Acq On : 15 Sep 2021 14:27
Operator : JC/MD
Sample : M3687-21
Misc : 5.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A005015

Quant Time: Sep 16 03:14:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

