

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068488.D
 Acq On : 15 Sep 2021 16:59
 Operator : JC/MD
 Sample : M3759-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31202

Quant Time: Sep 16 04:25:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	535495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	776171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	667924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	264904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	264566	53.860	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.720%			
35) Dibromofluoromethane	8.021	113	246123	52.576	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.160%			
50) Toluene-d8	10.446	98	904515	53.192	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.380%			
62) 4-Bromofluorobenzene	12.736	95	260071	47.796	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.600%			
Target Compounds					Qvalue	
25) 2-Butanone	7.351	43	4494	1.883	ug/l	95
40) Benzene	8.464	78	7578	0.396	ug/l	100
52) Toluene	10.507	92	8805	0.693	ug/l	96
68) m/p-Xylenes	11.953	106	6266	0.693	ug/l	93
69) o-Xylene	12.280	106	2354	0.267	ug/l	81
70) Styrene	12.296	104	4340	0.325	ug/l #	82
95) Naphthalene	15.512	128	11063	4.478	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
Data File : VN068488.D
Acq On : 15 Sep 2021 16:59
Operator : JC/MD
Sample : M3759-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31202

Quant Time: Sep 16 04:25:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

