

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068495.D
 Acq On : 15 Sep 2021 19:56
 Operator : JC/MD
 Sample : M3759-17 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31185

Quant Time: Sep 16 03:19:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	772332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1005404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	888558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	380949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	303930	42.900	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.800%
35) Dibromofluoromethane	8.024	113	297597	49.077	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.160%
50) Toluene-d8	10.443	98	1111729	50.472	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.940%
62) 4-Bromofluorobenzene	12.733	95	345866	49.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.140%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	23670	10.096	ug/l	94
52) Toluene	10.510	92	23141	1.406	ug/l	99
68) m/p-Xylenes	11.953	106	5238	0.435	ug/l	92
84) 1,2,4-Trimethylbenzene	13.369	105	5371	0.254	ug/l	82
95) Naphthalene	15.515	128	4784	3.791	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
Data File : VN068495.D
Acq On : 15 Sep 2021 19:56
Operator : JC/MD
Sample : M3759-17 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31185

Quant Time: Sep 16 03:19:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

