

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068475.D
 Acq On : 15 Sep 2021 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 15 16:36:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.408	0.390	4.4	96	0.00
3 P	Chloromethane	0.409	0.353	13.7	94	0.00
4 C	Vinyl Chloride	0.619	0.551	11.0#	92	0.00
5 T	Bromomethane	0.602	0.528	12.3	98	0.00
6 T	Chloroethane	0.456	0.404	11.4	94	0.00
7 T	Trichlorofluoromethane	0.776	0.706	9.0	95	0.00
8 T	Diethyl Ether	0.234	0.232	0.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.429	0.412	4.0	98	0.00
10 T	Methyl Iodide	0.612	0.620	-1.3	94	0.00
11 T	Tert butyl alcohol	0.074	0.060	18.9	91	0.00
12 CM	1,1-Dichloroethene	0.393	0.371	5.6#	93	0.00
13 T	Acrolein	0.034	0.025	26.5#	75	0.00
14 T	Allyl chloride	0.499	0.471	5.6	93	0.00
15 T	Acrylonitrile	0.175	0.167	4.6	94	0.00
16 T	Acetone	0.152	0.168	-10.5	124	0.00
17 T	Carbon Disulfide	1.055	0.966	8.4	93	0.00
18 T	Methyl Acetate	0.484	0.385	20.5	92	0.00
19 T	Methyl tert-butyl Ether	1.316	1.307	0.7	96	0.00
20 T	Methylene Chloride	0.522	0.442	15.3	98	0.00
21 T	trans-1,2-Dichloroethene	0.448	0.425	5.1	94	0.00
22 T	Diisopropyl ether	1.080	1.086	-0.6	95	0.00
23 T	Vinyl Acetate	0.807	0.830	-2.9	96	0.00
24 P	1,1-Dichloroethane	0.732	0.690	5.7	95	0.00
25 T	2-Butanone	0.223	0.227	-1.8	101	0.00
26 T	2,2-Dichloropropane	0.676	0.670	0.9	101	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.513	2.7	96	0.00
28 T	Bromochloromethane	0.285	0.279	2.1	89	0.00
29 T	Tetrahydrofuran	0.145	0.139	4.1	92	0.00
30 C	Chloroform	0.866	0.831	4.0#	94	0.00
31 T	Cyclohexane	0.695	0.622	10.5	91	0.00
32 T	1,1,1-Trichloroethane	0.803	0.792	1.4	94	0.00
33 S	1,2-Dichloroethane-d4	0.459	0.453	1.3	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.302	0.319	-5.6	96	0.00
36 T	1,1-Dichloropropene	0.417	0.426	-2.2	94	0.00
37 T	Ethyl Acetate	0.317	0.329	-3.8	95	0.00
38 T	Carbon Tetrachloride	0.499	0.542	-8.6	97	0.00
39 T	Methylcyclohexane	0.527	0.566	-7.4	95	0.00
40 TM	Benzene	1.234	1.288	-4.4	95	0.00
41 T	Methacrylonitrile	0.156	0.160	-2.6	95	0.00
42 TM	1,2-Dichloroethane	0.428	0.432	-0.9	95	0.00
43 T	Isopropyl Acetate	0.559	0.568	-1.6	94	0.00
44 TM	Trichloroethene	0.391	0.409	-4.6	95	0.00
45 C	1,2-Dichloropropane	0.295	0.306	-3.7#	96	0.00
46 T	Dibromomethane	0.237	0.249	-5.1	95	0.00
47 T	Bromodichloromethane	0.450	0.487	-8.2	97	0.00
48 T	Methyl methacrylate	0.234	0.238	-1.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
 Data File : VN068475.D
 Acq On : 15 Sep 2021 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 15 16:36:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	95	0.00
50 S	Toluene-d8	1.095	1.193	-8.9	95	0.00
51 T	4-Methyl-2-Pentanone	0.319	0.339	-6.3	95	0.00
52 CM	Toluene	0.819	0.912	-11.4#	97	0.00
53 T	t-1,3-Dichloropropene	0.431	0.501	-16.2	100	0.00
54 T	cis-1,3-Dichloropropene	0.472	0.533	-12.9	98	0.00
55 T	1,1,2-Trichloroethane	0.325	0.355	-9.2	96	0.00
56 T	Ethyl methacrylate	0.404	0.470	-16.3	96	0.00
57 T	1,3-Dichloropropane	0.501	0.546	-9.0	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.030	0.056	-86.7#	132	0.00
59 T	2-Hexanone	0.214	0.241	-12.6	99	0.00
60 T	Dibromochloromethane	0.382	0.459	-20.2	99	0.00
61 T	1,2-Dibromoethane	0.344	0.386	-12.2	96	0.00
62 S	4-Bromofluorobenzene	0.351	0.403	-14.8	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.443	0.472	-6.5	96	0.00
65 PM	Chlorobenzene	1.027	1.098	-6.9	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.404	0.447	-10.6	99	0.00
67 C	Ethyl Benzene	1.681	1.829	-8.8#	98	0.00
68 T	m/p-Xylenes	0.677	0.775	-14.5	100	0.00
69 T	o-Xylene	0.660	0.754	-14.2	101	0.00
70 T	Styrene	1.000	1.208	-20.8	102	0.00
71 P	Bromoform	0.310	0.389	-25.5#	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.557	3.530	0.8	99	0.00
74 T	N-amyl acetate	0.793	0.805	-1.5	99	0.00
75 P	1,1,2,2-Tetrachloroethane	0.997	0.912	8.5	99	0.00
76 T	1,2,3-Trichloropropane	0.789	0.713	9.6	107	0.00
77 T	Bromobenzene	0.984	1.009	-2.5	101	0.00
78 T	n-propylbenzene	3.756	3.790	-0.9	99	0.00
79 T	2-Chlorotoluene	2.315	2.253	2.7	101	0.00
80 T	1,3,5-Trimethylbenzene	2.889	2.966	-2.7	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.226	0.258	-14.2	109	0.00
82 T	4-Chlorotoluene	2.233	2.198	1.6	100	0.00
83 T	tert-Butylbenzene	2.638	2.730	-3.5	100	0.00
84 T	1,2,4-Trimethylbenzene	2.771	2.914	-5.2	101	0.00
85 T	sec-Butylbenzene	3.478	3.618	-4.0	100	0.00
86 T	p-Isopropyltoluene	2.936	3.173	-8.1	101	0.00
87 T	1,3-Dichlorobenzene	1.707	1.792	-5.0	101	0.00
88 T	1,4-Dichlorobenzene	1.682	1.742	-3.6	102	0.00
89 T	n-Butylbenzene	2.139	2.268	-6.0	101	0.00
90 T	Hexachloroethane	0.515	0.566	-9.9	106	0.00
91 T	1,2-Dichlorobenzene	1.647	1.725	-4.7	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.158	0.152	3.8	93	0.00
93 T	1,2,4-Trichlorobenzene	0.752	0.844	-12.2	93	0.00
94 T	Hexachlorobutadiene	0.486	0.535	-10.1	101	0.00
95 T	Naphthalene	1.536	1.650	-7.4	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091521\
Data File : VN068475.D
Acq On : 15 Sep 2021 10:21
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 15 16:36:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.718	0.768	-7.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6