

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091523\
 Data File : VN079276.D
 Acq On : 15 Sep 2023 16:54
 Operator : JC\MD
 Sample : 04408-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-16-0912

Quant Time: Sep 16 01:55:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	174463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	352935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	312569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	153652	65.243	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	130.480%#
35) Dibromofluoromethane	8.171	113	125133	53.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	10.571	98	421700	48.920	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.853	95	121648	44.241	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091523\
Data File : VN079276.D
Acq On : 15 Sep 2023 16:54
Operator : JC\MD
Sample : 04408-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-16-0912

Quant Time: Sep 16 01:55:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

