

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091525\
 Data File : VN087831.D
 Acq On : 15 Sep 2025 12:56
 Operator : JC\MD
 Sample : Q3082-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-32

Quant Time: Sep 16 01:35:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	167945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	378625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	385729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	171884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	179438	52.147	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.300%		
35) Dibromofluoromethane	8.153	113	142598	52.164	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.320%		
50) Toluene-d8	10.547	98	508535	49.702	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.400%		
62) 4-Bromofluorobenzene	12.829	95	175598	48.259	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 96.520%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	107284	57.274	ug/l	93
18) Methyl Acetate	5.030	43	20442	5.212	ug/l #	80
20) Methylene Chloride	5.259	84	11419	3.049	ug/l #	80
43) Isopropyl Acetate	8.677	43	20221	2.260	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091525\
Data File : VN087831.D
Acq On : 15 Sep 2025 12:56
Operator : JC\MD
Sample : Q3082-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-32

Quant Time: Sep 16 01:35:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

