

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058104.D
 Acq On : 16 Sep 2019 9:55
 Operator : JC/SP
 Sample : VN0916WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0916WBL01

Quant Time: Sep 17 04:00:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	329976	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	527896	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	475408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	249677	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	359257	53.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.50%	
35) Dibromofluoromethane	7.57	113	249157	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
50) Toluene-d8	10.09	98	953450	50.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.80%	
62) 4-Bromofluorobenzene	12.40	95	316596	43.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058104.D
Acq On : 16 Sep 2019 9:55
Operator : JC/SP
Sample : VN0916WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0916WBL01

Quant Time: Sep 17 04:00:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

