

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058108.D
 Acq On : 16 Sep 2019 11:22
 Operator : JC/SP
 Sample : K4871-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DAY-1

Manual Integrations
 APPROVED

MMDadoda
 9/18/2019 2:56:50 PM

Quant Time: Sep 17 09:40:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	297263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	497847	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	463154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	262427	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	341549	56.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.46%	
35) Dibromofluoromethane	7.58	113	233750	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	10.08	98	918057	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	12.40	95	319850	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.81	96	2727	0.595	ug/l	96
40) Benzene	8.03	78	14033	0.807	ug/l	94
78) n-propylbenzene	12.60	91	20196	0.864	ug/l	98
89) n-Butylbenzene	13.62	91	10003m	0.593	ug/l	
91) 1,2-Dichlorobenzene	13.66	146	40070	4.304	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058108.D
Acq On : 16 Sep 2019 11:22
Operator : JC/SP
Sample : K4871-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DAY-1

Manual Integrations
APPROVED

MMDadoda
9/18/2019 2:56:50 PM

Quant Time: Sep 17 09:40:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

