

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058110.D
 Acq On : 16 Sep 2019 12:06
 Operator : JC/SP
 Sample : K4877-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-4

Quant Time: Sep 17 09:01:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	360276	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	590618	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	539383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	366226	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
35) Dibromofluoromethane	7.58	113	254240	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
50) Toluene-d8	10.08	98	989656	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
62) 4-Bromofluorobenzene	12.40	95	333728	41.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.76%	
Target Compounds						
16) Acetone	3.80	43	18552	7.755	ug/l	96
20) Methylene Chloride	4.53	84	10815	1.173	ug/l	86
43) Isopropyl Acetate	8.15	43	103647	8.709	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058110.D
Acq On : 16 Sep 2019 12:06
Operator : JC/SP
Sample : K4877-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-4

Quant Time: Sep 17 09:01:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

