

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058112.D
 Acq On : 16 Sep 2019 12:51
 Operator : JC/SP
 Sample : K4661-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HD-02-091319

Quant Time: Sep 17 09:05:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	321975	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	538816	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	498576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	261153	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	355226	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
35) Dibromofluoromethane	7.58	113	245490	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.08	98	970406	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
62) 4-Bromofluorobenzene	12.40	95	325018	44.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.34%	
Target Compounds						
16) Acetone	3.80	43	13611	6.367	ug/l	98
20) Methylene Chloride	4.53	84	70249	16.435	ug/l	99
43) Isopropyl Acetate	8.15	43	115436	10.632	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058112.D
Acq On : 16 Sep 2019 12:51
Operator : JC/SP
Sample : K4661-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HD-02-091319

Quant Time: Sep 17 09:05:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

