

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058114.D
 Acq On : 16 Sep 2019 13:34
 Operator : JC/SP
 Sample : K4694-40
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-1-(0)

Quant Time: Sep 17 04:03:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	312857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	514865	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	473386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	251958	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	357952	56.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.98%	
35) Dibromofluoromethane	7.58	113	244916	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
50) Toluene-d8	10.09	98	963331	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	12.40	95	321023	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
Target Compounds						
16) Acetone	3.81	43	15575	7.498	ug/l	99
18) Methyl Acetate	4.32	43	5927	1.070	ug/l #	66
43) Isopropyl Acetate	8.15	43	107145	10.328	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058114.D
Acq On : 16 Sep 2019 13:34
Operator : JC/SP
Sample : K4694-40
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T3-1-(0)

Quant Time: Sep 17 04:03:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

