

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058119.D
 Acq On : 16 Sep 2019 15:23
 Operator : JC/SP
 Sample : K4719-30
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-3-(1.5)-HELLGATE

Quant Time: Sep 17 09:10:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	299317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	519314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	475662	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	250557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	362872	59.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.72%	
35) Dibromofluoromethane	7.57	113	249343	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
50) Toluene-d8	10.08	98	978550	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
62) 4-Bromofluorobenzene	12.41	95	326178	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
Target Compounds						
16) Acetone	3.81	43	16002	8.052	ug/l	97
20) Methylene Chloride	4.53	84	11646	1.894	ug/l	92
43) Isopropyl Acetate	8.16	43	141871	13.558	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058119.D
Acq On : 16 Sep 2019 15:23
Operator : JC/SP
Sample : K4719-30
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-3-(1.5)-HELLGATE

Quant Time: Sep 17 09:10:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

