

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058124.D
 Acq On : 16 Sep 2019 17:13
 Operator : JC/SP
 Sample : K4878-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-6

Quant Time: Sep 17 08:21:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	337954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	562879	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	520583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	275262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	357261	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	7.58	113	244321	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.08	98	930666	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
62) 4-Bromofluorobenzene	12.41	95	314826	40.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.92%	
Target Compounds						
16) Acetone	3.81	43	14242	6.347	ug/l	98
20) Methylene Chloride	4.53	84	121363	27.865	ug/l	98
43) Isopropyl Acetate	8.16	43	121067	10.674	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058124.D
Acq On : 16 Sep 2019 17:13
Operator : JC/SP
Sample : K4878-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-6

Quant Time: Sep 17 08:21:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

