

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058126.D
 Acq On : 16 Sep 2019 17:57
 Operator : JC/SP
 Sample : K4878-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-8

Quant Time: Sep 17 08:28:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	340345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	564706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	521290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	277641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	353202	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
35) Dibromofluoromethane	7.58	113	239282	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
50) Toluene-d8	10.09	98	951612	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
62) 4-Bromofluorobenzene	12.41	95	315093	40.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.72%	
Target Compounds						
16) Acetone	3.80	43	18456	8.167	ug/l	93
18) Methyl Acetate	4.30	43	24862	4.126	ug/l	97
20) Methylene Chloride	4.53	84	31742	6.303	ug/l	96
43) Isopropyl Acetate	8.15	43	90209	7.928	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058126.D
Acq On : 16 Sep 2019 17:57
Operator : JC/SP
Sample : K4878-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-8

Quant Time: Sep 17 08:28:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

