

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091619\
 Data File : VN058128.D
 Acq On : 16 Sep 2019 18:41
 Operator : JC/SP
 Sample : K4878-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TPA-9

Quant Time: Sep 17 08:33:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	350109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	587752	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	546864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	293597	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	357087	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
35) Dibromofluoromethane	7.58	113	248083	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
50) Toluene-d8	10.08	98	971390	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
62) 4-Bromofluorobenzene	12.41	95	326695	40.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.40%	
Target Compounds						
					Qvalue	
16) Acetone	3.80	43	14396	6.193	ug/l #	88
18) Methyl Acetate	4.31	43	7989	1.289	ug/l #	90
20) Methylene Chloride	4.53	84	24961	4.521	ug/l	97
43) Isopropyl Acetate	8.16	43	114808	9.694	ug/l #	86
95) Naphthalene	15.13	128	267986	14.476	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091619\
Data File : VN058128.D
Acq On : 16 Sep 2019 18:41
Operator : JC/SP
Sample : K4878-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TPA-9

Quant Time: Sep 17 08:33:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

