

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063367.D
 Acq On : 16 Sep 2020 16:19
 Operator : JC/MD
 Sample : PB131646TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646TB

Manual Integrations
 APPROVED

MMDadoda
 9/17/2020 1:08:53 PM

Quant Time: Sep 17 04:22:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	229213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	426433	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	416561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178512	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	187423	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
35) Dibromofluoromethane	7.55	113	138717	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	10.06	98	547367	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	12.38	95	220575	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
Target Compounds						
16) Acetone	3.78	43	13101	10.448	ug/l	96
18) Methyl Acetate	4.29	43	4701m	1.477	ug/l	
20) Methylene Chloride	4.51	84	11521	3.626	ug/l	94
43) Isopropyl Acetate	8.13	43	103791	13.018	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063367.D
Acq On : 16 Sep 2020 16:19
Operator : JC/MD
Sample : PB131646TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646TB

Manual Integrations
APPROVED

MMDadoda
9/17/2020 1:08:53 PM

Quant Time: Sep 17 04:22:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

