

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063375.D
 Acq On : 16 Sep 2020 19:30
 Operator : JC/MD
 Sample : PB131646ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646ZHE#08

Quant Time: Sep 17 04:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	249130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	458520	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	457989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	203375	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
35) Dibromofluoromethane	7.55	113	146738	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
50) Toluene-d8	10.06	98	595364	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.38	95	235312	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
Target Compounds						
16) Acetone	3.78	43	12598	9.243	ug/l #	86
18) Methyl Acetate	4.28	43	5184	1.498	ug/l #	70
20) Methylene Chloride	4.51	84	13748	3.981	ug/l	92
43) Isopropyl Acetate	8.13	43	91280	10.648	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063375.D
Acq On : 16 Sep 2020 19:30
Operator : JC/MD
Sample : PB131646ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646ZHE#08

Quant Time: Sep 17 04:36:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

