

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063377.D
 Acq On : 16 Sep 2020 20:18
 Operator : JC/MD
 Sample : PB131646ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646ZHE#10

Quant Time: Sep 17 04:39:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	245633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	465175	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	460450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	194793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	199053	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
35) Dibromofluoromethane	7.55	113	146579	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
50) Toluene-d8	10.06	98	595390	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	12.38	95	239109	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
Target Compounds						
16) Acetone	3.78	43	15020	11.177	ug/l	99
18) Methyl Acetate	4.29	43	5933	1.739	ug/l	93
20) Methylene Chloride	4.51	84	13418	3.941	ug/l	99
43) Isopropyl Acetate	8.13	43	89790	10.324	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063377.D
Acq On : 16 Sep 2020 20:18
Operator : JC/MD
Sample : PB131646ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646ZHE#10

Quant Time: Sep 17 04:39:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

