

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063378.D
 Acq On : 16 Sep 2020 20:42
 Operator : JC/MD
 Sample : PB131646ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646ZHE#11

Quant Time: Sep 17 04:41:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	232557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	436011	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431979	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	181822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	187040	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
35) Dibromofluoromethane	7.55	113	137877	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
50) Toluene-d8	10.06	98	564605	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	222409	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
Target Compounds						
16) Acetone	3.78	43	15072	11.847	ug/l	94
18) Methyl Acetate	4.29	43	4485	1.389	ug/l #	67
20) Methylene Chloride	4.50	84	15458	4.795	ug/l	85
43) Isopropyl Acetate	8.13	43	87507	10.735	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063378.D
Acq On : 16 Sep 2020 20:42
Operator : JC/MD
Sample : PB131646ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646ZHE#11

Quant Time: Sep 17 04:41:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

