

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063385.D
 Acq On : 17 Sep 2020 00:16
 Operator : JC/MD
 Sample : PB131646ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646ZHE#12

Quant Time: Sep 17 05:16:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	424763	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	419230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	173139	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	183235	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
35) Dibromofluoromethane	7.55	113	135271	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
50) Toluene-d8	10.06	98	547591	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	12.38	95	214116	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
Target Compounds						
16) Acetone	3.78	43	14520	11.722	ug/l	97
18) Methyl Acetate	4.29	43	5652	1.797	ug/l #	70
20) Methylene Chloride	4.50	84	15668	4.992	ug/l	98
43) Isopropyl Acetate	8.13	43	88937	11.199	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063385.D
Acq On : 17 Sep 2020 00:16
Operator : JC/MD
Sample : PB131646ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646ZHE#12

Quant Time: Sep 17 05:16:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

