

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063392.D
 Acq On : 17 Sep 2020 3:03
 Operator : JC/MD
 Sample : PB131646ZHE#19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131646ZHE#19

Quant Time: Sep 17 05:27:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	230125	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	429702	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	424935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	177577	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	184736	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
35) Dibromofluoromethane	7.55	113	137121	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
50) Toluene-d8	10.06	98	552812	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.38	95	224070	54.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.30%	
Target Compounds						
16) Acetone	3.78	43	14677	11.658	ug/l	99
18) Methyl Acetate	4.29	43	4317	1.351	ug/l	97
20) Methylene Chloride	4.51	84	14765	4.629	ug/l	93
43) Isopropyl Acetate	8.13	43	84530	10.522	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063392.D
Acq On : 17 Sep 2020 3:03
Operator : JC/MD
Sample : PB131646ZHE#19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131646ZHE#19

Quant Time: Sep 17 05:27:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

