

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063394.D
 Acq On : 17 Sep 2020 3:51
 Operator : JC/MD
 Sample : L3868-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-091520

Manual Integrations
 APPROVED

MMDadoda
 9/17/2020 1:44:15 PM

Quant Time: Sep 17 05:31:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	228021	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	432953	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	430062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	182975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	190584	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
35) Dibromofluoromethane	7.55	113	136613	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.16%	
50) Toluene-d8	10.06	98	554454	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	12.38	95	226312	54.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.56%	
Target Compounds						
16) Acetone	3.78	43	70682	56.661	ug/l	99
18) Methyl Acetate	4.29	43	4136m	1.306	ug/l	
20) Methylene Chloride	4.51	84	12580	3.980	ug/l	97
43) Isopropyl Acetate	8.13	43	70461	8.705	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063394.D
Acq On : 17 Sep 2020 3:51
Operator : JC/MD
Sample : L3868-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-091520

Manual Integrations
APPROVED

MMDadoda
9/17/2020 1:44:15 PM

Quant Time: Sep 17 05:31:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

