

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063395.D
 Acq On : 17 Sep 2020 4:15
 Operator : JC/MD
 Sample : L4015-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-6E-(A-0)

Quant Time: Sep 17 05:33:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	241151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	450376	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	447991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	196633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200267	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
35) Dibromofluoromethane	7.55	113	145735	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	10.06	98	586722	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	12.38	95	230331	53.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.18%	
Target Compounds						
16) Acetone	3.78	43	14810	11.226	ug/l	98
18) Methyl Acetate	4.28	43	5487	1.638	ug/l #	81
20) Methylene Chloride	4.50	84	12571	3.761	ug/l	90
43) Isopropyl Acetate	8.13	43	53808	6.390	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063395.D
Acq On : 17 Sep 2020 4:15
Operator : JC/MD
Sample : L4015-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-6E-(A-0)

Quant Time: Sep 17 05:33:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

