

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063396.D
 Acq On : 17 Sep 2020 4:39
 Operator : JC/MD
 Sample : L4015-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-6E-(B-0)

Quant Time: Sep 17 05:34:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	232235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	431365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	428789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	184538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194929	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
35) Dibromofluoromethane	7.55	113	142457	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
50) Toluene-d8	10.06	98	565370	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.38	95	222470	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
Target Compounds						
16) Acetone	3.78	43	14309	11.262	ug/l	98
18) Methyl Acetate	4.28	43	5912	1.833	ug/l #	72
20) Methylene Chloride	4.51	84	13626	4.233	ug/l	88
43) Isopropyl Acetate	8.13	43	57003	7.068	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063396.D
Acq On : 17 Sep 2020 4:39
Operator : JC/MD
Sample : L4015-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-6E-(B-0)

Quant Time: Sep 17 05:34:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

