

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063398.D
 Acq On : 17 Sep 2020 5:27
 Operator : JC/MD
 Sample : L4015-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-6E-(D-0)

Quant Time: Sep 17 06:43:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	230980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	443001	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	426947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	181099	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194765	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
35) Dibromofluoromethane	7.55	113	141758	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
50) Toluene-d8	10.06	98	567568	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
62) 4-Bromofluorobenzene	12.38	95	216377	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
Target Compounds						
16) Acetone	3.79	43	14550	11.514	ug/l	95
18) Methyl Acetate	4.29	43	4725	1.473	ug/l	99
20) Methylene Chloride	4.51	84	12987	4.056	ug/l	87
43) Isopropyl Acetate	8.13	43	79060	9.545	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063398.D
Acq On : 17 Sep 2020 5:27
Operator : JC/MD
Sample : L4015-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-6E-(D-0)

Quant Time: Sep 17 06:43:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

