

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091620\
 Data File : VN063403.D
 Acq On : 17 Sep 2020 7:27
 Operator : JC/MD
 Sample : L4018-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LACY-08

Quant Time: Sep 17 07:52:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	219381	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414162	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	405971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	176895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	184947	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	7.55	113	133612	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
50) Toluene-d8	10.06	98	531777	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.38	95	209664	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
Target Compounds						
16) Acetone	3.78	43	24602	20.498	ug/l	100
18) Methyl Acetate	4.28	43	4617	1.515	ug/l #	79
20) Methylene Chloride	4.51	84	12880	4.236	ug/l	95
43) Isopropyl Acetate	8.13	43	84486	10.911	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091620\
Data File : VN063403.D
Acq On : 17 Sep 2020 7:27
Operator : JC/MD
Sample : L4018-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LACY-08

Quant Time: Sep 17 07:52:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

