

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
 Data File : VN063379.D
 Acq On : 16 Sep 2020 21:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 04:43:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.566	0.534	5.7	97	0.00
3 P	Chloromethane	0.749	0.694	7.3	99	0.00
4 C	Vinyl Chloride	0.664	0.652	1.8#	99	0.00
5 T	Bromomethane	0.394	0.390	1.0	105	0.00
6 T	Chloroethane	0.408	0.377	7.6	100	0.00
7 T	Trichlorofluoromethane	0.991	0.953	3.8	98	0.00
8 T	Diethyl Ether	0.386	0.374	3.1	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.524	5.9	97	0.00
10 T	Methyl Iodide	0.725	0.696	4.0	96	0.00
11 T	Tert butyl alcohol	0.118	0.103	12.7	90	0.00
12 CM	1,1-Dichloroethene	0.564	0.532	5.7#	97	0.00
13 T	Acrolein	0.044	0.073	-65.9#	165#	0.00
14 T	Allyl chloride	1.116	1.062	4.8	95	0.00
15 T	Acrylonitrile	0.300	0.285	5.0	96	0.00
16 T	Acetone	0.274	0.250	8.8	98	0.00
17 T	Carbon Disulfide	1.641	1.509	8.0	94	0.00
18 T	Methyl Acetate	0.694	0.651	6.2	97	0.00
19 T	Methyl tert-butyl Ether	2.103	1.955	7.0	96	0.00
20 T	Methylene Chloride	0.693	0.635	8.4	99	0.00
21 T	trans-1,2-Dichloroethene	0.625	0.578	7.5	96	0.00
22 T	Diisopropyl ether	2.276	2.161	5.1	99	0.00
23 T	Vinyl Acetate	1.869	1.814	2.9	96	0.00
24 P	1,1-Dichloroethane	1.243	1.185	4.7	97	0.00
25 T	2-Butanone	0.421	0.389	7.6	94	0.00
26 T	2,2-Dichloropropane	1.094	0.924	15.5	88	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.677	4.4	96	0.00
28 T	Bromochloromethane	0.574	0.528	8.0	98	0.00
29 T	Tetrahydrofuran	0.287	0.263	8.4	96	0.00
30 C	Chloroform	1.247	1.193	4.3#	98	0.00
31 T	Cyclohexane	1.176	1.019	13.4	94	0.00
32 T	1,1,1-Trichloroethane	1.109	1.046	5.7	97	0.00
33 S	1,2-Dichloroethane-d4	0.851	0.758	10.9	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.346	0.300	13.3	95	0.00
36 T	1,1-Dichloropropene	0.533	0.488	8.4	96	0.00
37 T	Ethyl Acetate	0.523	0.480	8.2	95	0.00
38 T	Carbon Tetrachloride	0.544	0.499	8.3	95	0.00
39 T	Methylcyclohexane	0.578	0.525	9.2	95	0.00
40 TM	Benzene	1.521	1.413	7.1	98	0.00
41 T	Methacrylonitrile	0.235	0.239	-1.7	104	0.00
42 TM	1,2-Dichloroethane	0.612	0.565	7.7	96	0.00
43 T	Isopropyl Acetate	0.935	0.841	10.1	95	0.00
44 TM	Trichloroethene	0.369	0.339	8.1	98	0.00
45 C	1,2-Dichloropropane	0.414	0.397	4.1#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
 Data File : VN063379.D
 Acq On : 16 Sep 2020 21:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 04:43:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.272	0.252	7.4	96	0.00
47 T	Bromodichloromethane	0.565	0.540	4.4	99	0.00
48 T	Methyl methacrylate	0.459	0.419	8.7	95	0.00
49 T	1,4-Dioxane	0.007	0.006	14.3	94	0.00
50 S	Toluene-d8	1.302	1.150	11.7	95	0.00
51 T	4-Methyl-2-Pentanone	0.523	0.488	6.7	97	0.00
52 CM	Toluene	0.914	0.890	2.6#	99	0.00
53 T	t-1,3-Dichloropropene	0.631	0.592	6.2	97	0.00
54 T	cis-1,3-Dichloropropene	0.666	0.626	6.0	95	0.00
55 T	1,1,2-Trichloroethane	0.363	0.345	5.0	98	0.00
56 T	Ethyl methacrylate	0.567	0.547	3.5	99	0.00
57 T	1,3-Dichloropropane	0.655	0.615	6.1	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.284	4.4	102	0.00
59 T	2-Hexanone	0.379	0.357	5.8	97	0.00
60 T	Dibromochloromethane	0.410	0.394	3.9	99	0.00
61 T	1,2-Dibromoethane	0.377	0.361	4.2	99	0.00
62 S	4-Bromofluorobenzene	0.477	0.443	7.1	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.340	0.317	6.8	101	0.00
65 PM	Chlorobenzene	1.061	0.998	5.9	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.375	3.6	100	0.00
67 C	Ethyl Benzene	1.999	1.903	4.8#	100	0.00
68 T	m/p-Xylenes	0.721	0.685	5.0	99	0.00
69 T	o-Xylene	0.686	0.666	2.9	101	0.00
70 T	Styrene	1.151	1.150	0.1	100	0.00
71 P	Bromoform	0.290	0.284	2.1	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	4.007	3.770	5.9	99	0.00
74 T	N-amyl acetate	1.897	1.781	6.1	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.149	7.4	101	0.00
76 T	1,2,3-Trichloropropane	1.238	1.156	6.6	108	0.00
77 T	Bromobenzene	0.901	0.876	2.8	103	0.00
78 T	n-propylbenzene	4.667	4.414	5.4	100	0.00
79 T	2-Chlorotoluene	2.797	2.627	6.1	100	0.00
80 T	1,3,5-Trimethylbenzene	3.341	3.203	4.1	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.451	0.402	10.9	96	0.00
82 T	4-Chlorotoluene	2.956	2.781	5.9	101	0.00
83 T	tert-Butylbenzene	2.758	2.637	4.4	101	0.00
84 T	1,2,4-Trimethylbenzene	3.375	3.201	5.2	98	0.00
85 T	sec-Butylbenzene	3.691	3.503	5.1	100	0.00
86 T	p-Isopropyltoluene	3.256	3.098	4.9	99	0.00
87 T	1,3-Dichlorobenzene	1.671	1.564	6.4	99	0.00
88 T	1,4-Dichlorobenzene	1.698	1.578	7.1	98	0.00
89 T	n-Butylbenzene	3.001	2.796	6.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
Data File : VN063379.D
Acq On : 16 Sep 2020 21:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 17 04:43:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.620	0.583	6.0	99	0.00
91 T	1,2-Dichlorobenzene	1.640	1.545	5.8	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.262	11.2	93	0.00
93 T	1,2,4-Trichlorobenzene	0.967	0.866	10.4	90	0.00
94 T	Hexachlorobutadiene	0.400	0.350	12.5	91	0.00
95 T	Naphthalene	3.194	2.868	10.2	90	0.00
96 T	1,2,3-Trichlorobenzene	0.934	0.815	12.7	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6