

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
 Data File : VN063406.D
 Acq On : 17 Sep 2020 8:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 08:59:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.566	0.550	2.8	95	0.00
3 P	Chloromethane	0.749	0.717	4.3	97	0.00
4 C	Vinyl Chloride	0.664	0.692	-4.2#	100	0.00
5 T	Bromomethane	0.394	0.398	-1.0	102	0.00
6 T	Chloroethane	0.408	0.403	1.2	101	0.00
7 T	Trichlorofluoromethane	0.991	1.020	-2.9	100	0.00
8 T	Diethyl Ether	0.386	0.396	-2.6	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.528	5.2	92	0.00
10 T	Methyl Iodide	0.725	0.730	-0.7	96	0.00
11 T	Tert butyl alcohol	0.118	0.115	2.5	95	0.00
12 CM	1,1-Dichloroethene	0.564	0.556	1.4#	96	0.00
13 T	Acrolein	0.044	0.045	-2.3	96	0.00
14 T	Allyl chloride	1.116	0.955	14.4	81	0.00
15 T	Acrylonitrile	0.300	0.315	-5.0	101	0.00
16 T	Acetone	0.274	0.284	-3.6	105	0.00
17 T	Carbon Disulfide	1.641	1.584	3.5	93	0.00
18 T	Methyl Acetate	0.694	0.746	-7.5	106	0.00
19 T	Methyl tert-butyl Ether	2.103	2.082	1.0	97	0.00
20 T	Methylene Chloride	0.693	0.666	3.9	99	0.00
21 T	trans-1,2-Dichloroethene	0.625	0.608	2.7	95	0.00
22 T	Diisopropyl ether	2.276	2.261	0.7	98	0.00
23 T	Vinyl Acetate	1.869	1.907	-2.0	96	0.00
24 P	1,1-Dichloroethane	1.243	1.236	0.6	96	0.00
25 T	2-Butanone	0.421	0.439	-4.3	101	0.00
26 T	2,2-Dichloropropane	1.094	0.687	37.2#	62	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.706	0.3	95	0.00
28 T	Bromochloromethane	0.574	0.560	2.4	98	0.00
29 T	Tetrahydrofuran	0.287	0.293	-2.1	101	0.00
30 C	Chloroform	1.247	1.242	0.4#	97	0.00
31 T	Cyclohexane	1.176	1.063	9.6	93	0.00
32 T	1,1,1-Trichloroethane	1.109	1.101	0.7	97	0.00
33 S	1,2-Dichloroethane-d4	0.851	0.849	0.2	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.346	0.317	8.4	99	0.00
36 T	1,1-Dichloropropene	0.533	0.492	7.7	95	0.00
37 T	Ethyl Acetate	0.523	0.518	1.0	100	0.00
38 T	Carbon Tetrachloride	0.544	0.517	5.0	96	0.00
39 T	Methylcyclohexane	0.578	0.507	12.3	90	0.00
40 TM	Benzene	1.521	1.441	5.3	97	0.00
41 T	Methacrylonitrile	0.235	0.233	0.9	100	0.00
42 TM	1,2-Dichloroethane	0.612	0.584	4.6	97	0.00
43 T	Isopropyl Acetate	0.935	0.897	4.1	99	0.00
44 TM	Trichloroethene	0.369	0.349	5.4	98	0.00
45 C	1,2-Dichloropropane	0.414	0.400	3.4#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
 Data File : VN063406.D
 Acq On : 17 Sep 2020 8:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 17 08:59:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.272	0.267	1.8	100	0.00
47 T	Bromodichloromethane	0.565	0.554	1.9	100	0.00
48 T	Methyl methacrylate	0.459	0.450	2.0	100	0.00
49 T	1,4-Dioxane	0.007	0.006	14.3	100	0.00
50 S	Toluene-d8	1.302	1.238	4.9	100	0.00
51 T	4-Methyl-2-Pentanone	0.523	0.530	-1.3	103	0.00
52 CM	Toluene	0.914	0.905	1.0#	99	0.00
53 T	t-1,3-Dichloropropene	0.631	0.577	8.6	92	0.00
54 T	cis-1,3-Dichloropropene	0.666	0.602	9.6	90	0.00
55 T	1,1,2-Trichloroethane	0.363	0.357	1.7	99	0.00
56 T	Ethyl methacrylate	0.567	0.576	-1.6	102	0.00
57 T	1,3-Dichloropropane	0.655	0.629	4.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.280	5.7	99	0.00
59 T	2-Hexanone	0.379	0.384	-1.3	102	0.00
60 T	Dibromochloromethane	0.410	0.401	2.2	99	0.00
61 T	1,2-Dibromoethane	0.377	0.375	0.5	100	0.00
62 S	4-Bromofluorobenzene	0.477	0.468	1.9	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.340	0.330	2.9	103	0.00
65 PM	Chlorobenzene	1.061	1.012	4.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.378	2.8	100	0.00
67 C	Ethyl Benzene	1.999	1.904	4.8#	99	0.00
68 T	m/p-Xylenes	0.721	0.693	3.9	99	0.00
69 T	o-Xylene	0.686	0.667	2.8	100	0.00
70 T	Styrene	1.151	1.163	-1.0	100	0.00
71 P	Bromoform	0.290	0.286	1.4	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.007	3.882	3.1	99	0.00
74 T	N-amyl acetate	1.897	1.873	1.3	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.202	3.1	101	0.00
76 T	1,2,3-Trichloropropane	1.238	1.120	9.5	101	0.00
77 T	Bromobenzene	0.901	0.893	0.9	101	0.00
78 T	n-propylbenzene	4.667	4.474	4.1	98	0.00
79 T	2-Chlorotoluene	2.797	2.731	2.4	101	0.00
80 T	1,3,5-Trimethylbenzene	3.341	3.227	3.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.451	0.376	16.6	87	0.00
82 T	4-Chlorotoluene	2.956	2.819	4.6	99	0.00
83 T	tert-Butylbenzene	2.758	2.674	3.0	99	0.00
84 T	1,2,4-Trimethylbenzene	3.375	3.279	2.8	97	0.00
85 T	sec-Butylbenzene	3.691	3.543	4.0	98	0.00
86 T	p-Isopropyltoluene	3.256	3.099	4.8	95	0.00
87 T	1,3-Dichlorobenzene	1.671	1.612	3.5	98	0.00
88 T	1,4-Dichlorobenzene	1.698	1.642	3.3	98	0.00
89 T	n-Butylbenzene	3.001	2.764	7.9	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
Data File : VN063406.D
Acq On : 17 Sep 2020 8:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 17 08:59:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.620	0.590	4.8	97	0.00
91 T	1,2-Dichlorobenzene	1.640	1.579	3.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.283	4.1	97	0.00
93 T	1,2,4-Trichlorobenzene	0.967	0.868	10.2	87	0.00
94 T	Hexachlorobutadiene	0.400	0.357	10.8	89	0.00
95 T	Naphthalene	3.194	3.030	5.1	92	0.00
96 T	1,2,3-Trichlorobenzene	0.934	0.857	8.2	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6