


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091620\
 Data File : VN063356.D
 Acq On : 16 Sep 2020 10:05
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Title : SW846 8260
 Last Update : Wed Sep 16 13:05:20 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	12429	PASS
75	95	30	60	57.0	30037	PASS
95	95	100	100	100.0	52685	PASS
96	95	5	9	6.8	3581	PASS
173	174	0.00	2	1.9	728	PASS
174	95	50	100	71.8	37837	PASS
175	174	5	9	7.6	2861	PASS
176	174	95	101	96.8	36613	PASS
177	176	5	9	6.2	2285	PASS