

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091621\
 Data File : VN068500.D
 Acq On : 16 Sep 2021 16:19
 Operator : JC/MD
 Sample : VN0916WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0916WBL01

Quant Time: Sep 17 02:01:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	554582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	793554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	688968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	269984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	265200	52.131	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.260%	
35) Dibromofluoromethane	8.024	113	247502	51.712	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.420%	
50) Toluene-d8	10.443	98	924286	53.164	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.320%	
62) 4-Bromofluorobenzene	12.736	95	271151	48.741	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091621\
Data File : VN068500.D
Acq On : 16 Sep 2021 16:19
Operator : JC/MD
Sample : VN0916WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0916WBL01

Quant Time: Sep 17 02:01:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

