

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091621\
 Data File : VN068508.D
 Acq On : 16 Sep 2021 20:51
 Operator : JC/MD
 Sample : M3769-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1670-1675

Manual Integrations
 APPROVED

MMDadoda
 9/17/2021 2:33:01 PM

Quant Time: Sep 17 02:03:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	514304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	748308	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	667434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	290609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	253966	53.832	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.660%	
35) Dibromofluoromethane	8.027	113	230052	50.972	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.940%	
50) Toluene-d8	10.446	98	869303	53.025	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.040%	
62) 4-Bromofluorobenzene	12.734	95	271222	51.701	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	6088m	3.900	ug/l	
84) 1,2,4-Trimethylbenzene	13.369	105	58531	3.634	ug/l	96
95) Naphthalene	15.509	128	26045	5.602	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091621\
Data File : VN068508.D
Acq On : 16 Sep 2021 20:51
Operator : JC/MD
Sample : M3769-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1670-1675

Manual Integrations
APPROVED

MMDadoda
9/17/2021 2:33:01 PM

Quant Time: Sep 17 02:03:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

