

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
 Data File : VN074503.D
 Acq On : 16 Sep 2022 12:19
 Operator : JC\MD
 Sample : N4626-09RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-418-420RE

Quant Time: Sep 16 16:14:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	338903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	649759	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	626359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	268451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	263951	45.902	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.800%
35) Dibromofluoromethane	7.957	113	216848	50.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.480%
50) Toluene-d8	10.380	98	788225	47.220	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.440%
62) 4-Bromofluorobenzene	12.674	95	300181	52.506	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.020%
Target Compounds						
16) Acetone	4.234	43	12009	3.264	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
Data File : VN074503.D
Acq On : 16 Sep 2022 12:19
Operator : JC\MD
Sample : N4626-09RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-418-420RE

Quant Time: Sep 16 16:14:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

