

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
 Data File : VN074513.D
 Acq On : 16 Sep 2022 16:22
 Operator : JC\MD
 Sample : N4508-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-090222

Quant Time: Sep 16 22:33:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	339938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	648576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	633314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	261095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	262166	45.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.900%
35) Dibromofluoromethane	7.951	113	216787	50.318	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.640%
50) Toluene-d8	10.380	98	785663	47.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.674	95	299627	52.504	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.000%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	32017	8.675	ug/l #	88
20) Methylene Chloride	5.022	84	4409	0.813	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
Data File : VN074513.D
Acq On : 16 Sep 2022 16:22
Operator : JC\MD
Sample : N4508-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-090222

Quant Time: Sep 16 22:33:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

