

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
 Data File : VN074514.D
 Acq On : 16 Sep 2022 16:46
 Operator : JC\MD
 Sample : N4508-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB02-090222

Quant Time: Sep 16 22:34:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	311947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	630003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	620443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	262421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	258669	48.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.740%
35) Dibromofluoromethane	7.951	113	208113	49.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.460%
50) Toluene-d8	10.381	98	775713	47.928	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.674	95	292144	52.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.400%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	36322	10.724	ug/l #	75
25) 2-Butanone	7.263	43	10724	1.907	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
Data File : VN074514.D
Acq On : 16 Sep 2022 16:46
Operator : JC\MD
Sample : N4508-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB02-090222

Quant Time: Sep 16 22:34:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

